



Maine EMS Approved Equipment

(as of May 2011)

- 1) CPAP Device Requirements:
 1. Full Face Mask
 2. Continuous flow device
 3. Must be a device capable of adjusting FiO₂ so must run on Oxygen
 4. Capable of regulating PEEP (7.5 to 10cm H₂O)
 5. Must be completely latex-free.
 - 6 The ability to attach a nebulizer is optional but encouraged

- 2) EZ-IO (tibial and humeral head in adult, tibial in peds)
- 3) FAST1 sternal IO- LOM only
- 4) 11/15/06- Masimo Rad 57 Carbon Monoxide Detector- Rockland FD
- 5) 09/08/08 – Turkel needle for chest decompression only
- 6) 10/10/08 – SureVent from Hartwell Medical *NOTE: for pre-hospital cardiac arrest use only; not for transfers
- 7) 11/26/08- X-Collar, the cervical splint
- 8) 07/07/09 – Cook Needle Decompression Kit
- 9) 06/17/09 – 3.25 needle for chest decompression only
- 10) 08/31/09 – O₂-RESQ with BiTrac ED Mask (disposable CPAP) from BoundTree Medical
- 11) 06/16/10 - S-SCORT S3 oropharyngeal evacuation tool as an optional additional device
- 12) Pertrach, Quick Trach, Nu-trake, etc
Note MDPB approved cric kits only if they follow the method of piecing the cricothyroid membrane